

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA SUB 5x1.5mL

REF      5006235  
LOT      8F41B00.03  
      31-Mai-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1.5mL	TGA Substrat	8F41B00	31-Mai-2016


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F31B00	276 - 512	441.5	OK
		TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D41B00	190 - 352	305.1	OK

Durchführung von  
Prepared by

**Fikret NASUFI**  


04-Dec-2014  
Datum / Date:

Freigabe von  
Approved by

  
**Lavdrim NASUFI**

04-Dec-2014  
Datum / Date: