

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA SUB 5x1.5mL

REF 5006235
LOT 8F41B01.01
 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1.5mL	TGA Substrat	8F41B01	30-Sep-2017

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	TGA RC High LOT 7F51B00	276 - 512	420.7	OK
		TGA RC Low LOT 7D51B00	190 - 352	258	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

28-Sep-2015
Datum / Date:

Freigabe von
Approved by

Michael KAFKA
Michael Kafka

28-Sep-2015
Datum / Date: