

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8E00AZ
---	---	----------

Ceveron® TGA SUB 5x3 mL

REF 5006237
LOT 8E11B01.01
 30 - Nov - 2015

Inhalt Content	Reagenz Reagent	LOT	
5x3 mL	Ceveron® TGA SUB	8E11B01	30-Nov-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	195.3 – 362.7	268.8	OK
	TGA Control High (Lag time min) LOT 1B12B01)	1.47 – 2.73	2.4	OK
	TGA Control High (AUC) LOT 1B12B01)	1985.9 – 3688.1	2971.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUF 

19-FEB-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA 

14-FEB-2014
Datum / Date: