



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8E00AZ

Ceveron[®] TGA SUB 5x3 mL

REF 5006237
LOT 8E11B01.02
⌚ 30 - Nov - 2015

Inhalt Content	Reagenz Reagent	LOT	⌚
5x3 mL	Ceveron [®] TGA SUB	8E11B01	30-Nov-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung RC Low LOT 7F41B	Peak Thrombin nM)	TGA Control High (nM Peak Thrombin) LOT 1B41B00	188 - 350	268.1	OK
		TGA Control Low (nM Peak Thrombin) LOT 2B43B00	47 - 87	67.3	OK
Recovery of controls RC Low LOT 7F41B	AUC nM	TGA Control High (nM Peak Thrombin) LOT 1B41B00	2082 - 3866	2749.0	OK
		TGA Control Low (nM Peak Thrombin) LOT 2B43B00	527 - 979	696.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

03-Oct-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

03-Oct-2014
Datum / Date: