

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1B00AZ

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320
LOT 1B12B01.01
 30 – Apr - 2015

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1B12B01	30-Apr-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F23B00	316 - 586 473.8 OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D12B00	231 - 429 352.9 OK
	TGA Reagenz RD (nm Peak Thrombin) LOT 7G83A03	166 - 308 281.0 OK

Durchführung von
Prepared by

J. Nasozi

14. Juni 2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK
[Signature]

14. JUNI 2013
Datum / Date: