

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1B00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320
LOT 1B41B00.01
 30 - Apr - 2016

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1B41B00	30-Apr-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F42B00	276 - 512	350.9	OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D41B00	190 - 352	229.9	OK
	TGA Reagenz RD (nm Peak Thrombin) LOT 7G31B00	161 - 299	222.9	OK
	TGA Reagenz RB (nm Peak Thrombin) LOT 7C31B00	254 - 472	405.6	OK

**Durchführung von
Prepared by**

Fikret NASUFI
Fikret Nasufi

01-Jul-2014
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA
Michael Riha

01-Jul-2014
Datum / Date: