

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1B00AZ
---	---	----------

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320
LOT 1B41B00.02
 30-Apr-2016

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1B41B00	30-Apr-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F51B00	276 - 512	431.2 OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D51B00	190 - 352	286.2 OK
	TGA Reagenz RD (nm Peak Thrombin) LOT 7G51B00	161 - 299	253.6 OK
	TGA Reagenz RB (nm Peak Thrombin) LOT 7C51B00	254 - 472	313.7 OK

**Durchführung von
Prepared by**

Fikret NASUFI
Fikret Nasufi

28-Sep-2015
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA
[Signature]

28-Sep-2015
Datum / Date: