
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ1B00AZ.03
---	---	-------------

### Ceveron Technothrombin TGA Control High 5x1ml

REF      5006322  
LOT      1B12B01.01  
      30-Apr-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	RC Low : Lag Time	2.4 – 4.4 min	3.45 min	OK
	RC Low : Peak Thrombin	142 – 264 nM	178.2 nM	OK
	RC Low : AUC	2023 - 3757	2905.55	OK

**Durchführung von**  
**Prepared by**

**Alessandro BENIGNI**

\_\_\_\_\_

*21-Jun-2013*  
Datum / Date:

**Freigabe von**  
**Approved by**

**Harald SEVCIK**

\_\_\_\_\_

21. JUNI 2013  
Datum / Date: