




# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ1B00AZ.02

## Ceveron TGA Control High 5x1mL

REF 5006322

LOT 1B41B00.01

 30-Apr-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		7.4 – 7.8	7.7	OK	
Verwendbarkeit: Usability: Ceveron® alpha	Reagent B 7C31B	Peak Thrombin (nM)	120 - 222	199.0	OK
	Reagent RC Low 7D32B	Peak Thrombin (nM)	106 - 198	186.7	OK
	Reagent RC High 7F41B	Peak Thrombin (nM)	188 - 350	299.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

**Fikret NASUFI**

*Fikret Nasufi*

01-Jul-2014  
Datum / Date:

Freigabe von  
Approved by

**Michael RIHA**

*Michael Riha*

01-Jul-2014  
Datum / Date: