



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1B00AZ.02

Ceveron TGA Control High 5x1mL

REF 5006322
LOT 1B41B00.02
⌚ 30-Apr-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		7.4 – 7.8	7.7	OK	
Verwendbarkeit: Usability: Ceveron® alpha	Reagent RC Low 7D42B	Peak Thrombin (nM)	106 - 198	182.2	OK
	Reagent RC High 7F41B	Peak Thrombin (nM)	188 - 350	260.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Fikret NASUFI

29-Jul-2015
Datum / Date:

Freigabe von
Approved by

[Signature]

28-Jul-2015
Datum / Date: