

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ2B00AZ
---	---	----------

## TECHNOTHROMBIN® TGA Control Low 5x1mL

REF      5006330  
LOT      2B43B00.01  
      30 - Jun - 2016

<b>Inhalt</b> <b>Content</b>	<b>Reagenz</b> <b>Reagent</b>	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1mL	TGA Control Low	2B43B00	30-Jun-2016

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>	<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
		Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F42B00
TGA Reagenz RC Low (nm Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D41B00	48 - 88		51.5    OK
TGA Reagenz RD (nm Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7G31B00	239 - 443		244.2    OK

**Durchführung von**  
**Prepared by**

**Fikret NASUFI** *Fikret Nasufi*

*01-Jul-2014*  
Datum / Date:

**Freigabe von**  
**Approved by**

**Michael RIHA** *Michael Riha*

*01-Jul-2014*  
Datum / Date: