



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ2B00AZ.02

Ceveron TGA Control Low 5x1ml

REF 5006332
 LOT 2B34B00.02
 ⌚ 30-Jun-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.6	OK

Verwendbarkeit Usability Ceveron® alpha	Reagent B 7C31B	Peak Thrombin (nM)	29 - 55	35.8	OK
		Lag time (min)	13.3 - 24.7	13.9	OK
		AUC (nM)	507 - 941	746.2	OK
	Reagent RC Low 7D32B	Peak Thrombin (nM)	23 - 43	31.3	OK
		Lag time (min)	7.7 - 14.3	9.6	OK
		AUC (nM)	437 - 811	663.3	OK
	Reagent RC High 7F41B	Peak Thrombin (nM)	61.88 - 114.9	70.9	OK
		Lag time (min)	3.01 - 5.59	4.6	OK
		AUC (nM)	623.7 – 1158.3	886.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von
Prepared by

Fikret NASUFI

Fikret Nasufi

22-Mai-2014
Datum / Date:

Freigabe von
Approved by

[Signature]

22-Mai-2014
Datum / Date: