



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2B00AZ.02

Ceveron TGA Control Low 5x1ml

REF 5006332

LOT 2B43B00.01

🕒 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		7.4 – 7.8	7.7	OK	
Verwendbarkeit Usability Ceveron® alpha	Reagent B 7C31B	Peak Thrombin (nM)	21 - 39	37.8	OK
	Reagent RC Low 7D32B	Peak Thrombin (nM)	23 - 43	29.5	OK
	Reagent RC High 7F41B	Peak Thrombin (nM)	47 - 87	85.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

01-Jul-2016
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

01-Jul-2016
Datum / Date: