



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ5T00AZ

TECHNOTHROMBIN® TGA Calibrator Set

REF 50 063 45

LOT 0D12B00.03

⌚ 30-Jun-2015

Inhalt /  
Content

Reagenz / reagent

LOT



1 x 3 mL

TGA buffer

9P01B01

31-Dec-2015

1 x 0.5mL

TGA Thrombin Calibrator

5T13B00

30-Jun-2015

Test	Spezifikation / specification	Ergebnis / result
Kalibrierung / Calibration gegen / against WHO 01/580	800-900 nM	841 nM Thrombin
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader	Regression der Bezugskurve / regression of the calibration curve ≥ 0.990	0.995

Erstellt / created: F.Nasufi

Unterschrift / signature: *F. Nasufi*

Datum / date: 24-Mai-2013

Charge freigegeben / batch is released:  Ja / yes

Nein / no

Freigegeben / released: Harald SEVCIK

Unterschrift / signature:

Datum / date: 27. MAI 2013