

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ5T00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Cal Set

REF 5006345
LOT 0D41B00.01
 30-Jun-2015

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P31B00	30-Sep-2015
1 x 0.5mL	TGA Thrombin Calibrator	5T13B	30-Jun-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580	800-900 nM	841	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve ≥ 0.990	0.998	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

07-Nov-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

07-Nov-2014
Datum / Date: