

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ5T00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA Cal Set

REF      5006345  
LOT      0D43B00.01  
      31-Dec-2015

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 3 mL	TGA buffer	9P01B01	31-Dec-2015
1 x 0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		800-900 nM	841	OK
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.991	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

15-Jan-2015

Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI

19-Jan-2015

Datum / Date: