

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: right;">DQ5T00AZ</p>
---	---	--

TECHNOTHROMBIN® TGA Cal Set

REF 5006345
LOT 0D53B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P01B02	31-Jul-2017
1 x 0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		800-900 nM	841	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.994	OK

Durchführung von
Prepared by

Fikret MASUFI
Fikret Masufi

08-Sep-2015
Datum / Date:

Freigabe von
Approved by



08-SEP-2015
Datum / Date: