



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0D00AZ

Ceveron TGA CAL Set f. Ceveron

REF 5006347  
LOT 0D31B00.01  
⌚ 30-Jun-2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1x3mL	TGA Buffer	9P31B00	30-Sep-2015
1x1mL	TGA Reaction Buffer	9K31B	30-Nov-2015
1x0.5mL	TGA Thrombin Standard	5T13B00	30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Verwendbarkeit: Usability: Reagent B Ceveron® alpha	Technothrombin® TGA Control High LOT 1B12B01	Lag time	5.5 – 10.3 min	7.1	OK
		Peak thrombin	153 – 283 nM	230	OK
		AUC	1894 – 3517 nM	3037	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Lavdrim NASUFI

11-Dec-2013  
Datum / Date:

Freigabe von  
Approved by

Lucas HOLOVICZ

11-Dec-2013  
Datum / Date: