

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0D00AZ
---	---	----------

## CEVERON TGA Cal Set f.Ceveron

REF      5006347  
LOT        0D51B00.01  
      30-Nov-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 3 mL	TGA buffer	9P31B01	30-Apr-2017
1 x 0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016
1 x 1 mL	TGA Reaction Buffer f.Ceveron	9K51B	30-Apr-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580	800-900 nM	841	OK
Verwendbarkeit / usability  Gerät / Analyzer: Ceveron:106	Regression der Bezugskurve / regression of the calibration curve  ≥ 0.990	1.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK

Durchführung von  
Prepared by

Fikret NASUFI 

07-May-2015  
Datum / Date:


Freigabe von  
Approved by

Michael RIHA 

07-MAR-2015  
Datum / Date:

Ceveron® alpha

n.z.

S.SCHWINGENSCHLÖGL 

07-MAR-2015  
Datum / Date: