

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ2B00AZ.01
---	---	-------------

## TECHNOTHROMBIN<sup>®</sup> TGA MFC 5x1 mL

REF      5006351  
LOT      2B510RW.01  
      29-Feb-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5 x 1 mL	TGA MFC 1mL	2B510RW	29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	TGA MFC (2B510RW)	PPP: (nM Peak Thrombin) 31-46	42.3	OK
		MFP: (nM Peak Thrombin) 26-39	29.4	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

25-Mar-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael RIHA *Michael Riha*

25-MAR-2015  
Datum / Date: