
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1M00AZ.01
---	---	-------------

TECHNOTHROMBIN[®] TGA MHC 5x1 mL

REF 5006352
LOT 1M51B00.01
 29-Feb-2016

Inhalt Content	Reagenz Reagent	LOT	
5 x 1 mL	TGA MHC 1mL	1M51B00	29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	TGA MHC 1mL (1M51B00)	PPP: (nM Peak Thrombin) 169-253	182.4	OK
		MFP: (nM Peak Thrombin) 37-55	39.0	OK

Durchführung von
Prepared by

Fikret NASUFI 

25-Nov-2015
Datum / Date:

Freigabe von
Approved by

Michael RIHA 

25-NOV-2015
Datum / Date: