

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9K00AZ
---	---	----------

## Ceveron<sup>®</sup> TGA Reaction BUF [2] 5x1 mL

REF      5006370  
LOT      9K31B.02  
      30 - Nov - 2015

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1 mL	Ceveron <sup>®</sup> TGA Reaction BUF [2]	9K31B	30-Nov-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B12B01)	195.3 – 362.7	268.8	OK
	TGA Control High (Lag time min) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B12B01)	1.47 – 2.73	2.4	OK
	TGA Control High (AUC) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B12B01)	1985.9 – 3688.1	2971.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

**FIKRET NASUFI**  
*Fikret Nasufi*

*19-Feb-2014*  
Datum / Date:

Freigabe von  
Approved by

**Michael RIHA**  
*Michael Riha*

*19-Feb-2014*  
Datum / Date: