
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9K00AZ
---	---	----------

### Ceveron TGA Reaction BUF [2] 5x1mL

**REF**            5006370  
**LOT**            9K31B.03  
        30-Nov-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		7.0 – 7.6	7.2	OK	
Verwendbarkeit: Usability: <b>Reagent B</b> Ceveron <sup>®</sup> alpha	Technothrombin <sup>®</sup> TGA Control High <b>LOT</b> 1B12B01	Lag time	5.5 – 10.3 min	8.2	OK
		Peak thrombin	153 – 283 nM	242.8	OK
		AUC	1894 – 3517 nM	3152.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

**Durchführung von**  
**Prepared by**

**Fikret NASUFI**  


*11-Jun-2014*  
Datum / Date:

**Freigabe von**  
**Approved by**

**Michael RIHA**  


*11-Jun-2014*  
Datum / Date: