

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9K00AZ
---	---	----------

## Ceveron® TGA Reaction BUF [2] 5x1 mL

REF      5006370  
LOT      9K51B.01  
      30-Apr-2017

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1 mL	Ceveron® TGA Reaction BUF [2]	9K51B	30-Apr-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00)	188 – 350	215.7	OK
	TGA Control High (Lag time min) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00)	1.5 – 2.7	2.3	OK
	TGA Control High (AUC) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00)	2082 – 3866	2974.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von  
Prepared by

**Fikret NASUFI**

29-Jul-2015  
Datum / Date:

Freigabe von  
Approved by



24-Jul-2015  
Datum / Date: