

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: right;">DQ9K00AZ</p>
---	---	--

Ceveron[®] TGA Reaction BUF [2] 5x1 mL

REF 5006370
LOT 9K51B.02
 30-Apr-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron [®] TGA Reaction BUF [2]	9K51B	30-Apr-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	188 – 350	304.2	OK
	TGA Control High (Lag time min) LOT 1B41B00)	1.5 – 2.7	2.1	OK
	TGA Control High (AUC) LOT 1B41B00)	2082 – 3866	3183.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI

06-Oct-2015
Datum / Date:

Freigabe von
Approved by

Michael KAFKA

06-Oct-2015
Datum / Date: