


AK Calibrant 4x1 mL

REF 5010004
LOT 0U42B00.01
⌚ 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	AK Calibrant A	1N23B03	31-Oct-2018
1x1mL	AK Calibrant B	3L34B01	31-Oct-2018
1x1mL	AK Calibrant C	3M34B01	31-Oct-2018
1x1mL	AK Calibrant D	3N34B01	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P44B00 Recombiplastin 2G %	92.9 – 117.1 %	104.4	OK
	CCN LOT 1P44B00 Recombiplastin 2G INR	0.86 – 1.08	0.97	OK
	CCA LOT 3P38B00 Recombiplastin 2G %	29.2 – 36.8 %	32.5	OK
	CCA LOT 3P38B00 Recombiplastin 2G INR	2.05 – 2.59	2.31	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lavdrim NASUFI


18-Dec-2014
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

18-Dec-2014
Datum / Date: