



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0U00AZ.02

## AK Calibrant 4x1 mL

**REF** 5010004  
**LOT** 0U42B00.02  
 31-Oct-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x1mL	AK Calibrant A	1N23B03	31-Oct-2018
1x1mL	AK Calibrant B	3L34B01	31-Oct-2018
1x1mL	AK Calibrant C	3M34B01	31-Oct-2018
1x1mL	AK Calibrant D	3N34B01	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TCN <b>LOT</b> 1A31B02 Technoplastin HIS %	89.1 – 112.3 %	105.3	OK
	TCN <b>LOT</b> 1A31B02 Technoplastin HIS INR	0.87 – 1.09	0.95	OK
	CCA <b>LOT</b> 3P38B00 Technoplastin HIS %	25.7 – 32.3 %	29.3	OK
	CCA <b>LOT</b> 3P38B00 Technoplastin HIS INR	2.54 – 3.20	2.86	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lucas Rosenmayr

02-Mär-2015  
Datum / Date:

Freigabe von  
Approved by

Laydrim NASUFI

02-Mär-2015  
Datum / Date: