





**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ0U00AZ.02

**AK Calibrant**

**REF**            5010004  
**LOT**            0U31B02.01  
            30-Jun-2016

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x1mL	AK Calibrant A	1N23B02	30-Jun-2016
1x1mL	AK Calibrant B	3L26B02	30-Jun-2016
1x1mL	AK Calibrant C	3M25B02	31-Aug-2016
1x1mL	AK Calibrant D	3N2AB02	31-Aug-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TCN <b>LOT</b> 1A31B01 Technoplastin HIS %	89.1 – 112.3%	98.9	OK
	TCA <b>LOT</b> 3P31B01 Technoplastin HIS %	29.2 – 36.8%	31.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

Alessandro BENIGNI

04-Jun-2014  
Datum / Date:

**Freigabe von  
Approved by**

Lucas HOLOVICZ

04-Jun-2014  
Datum / Date: