

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ.02
---	---	-------------

Ceveron® Coagulation Control AK 5x1mL

REF	5011030
LOT	2B35B00.03
	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	TECHNOPLASTIN® HIS (%)	30.8 – 38.8	31.2	OK
	TECHNOPLASTIN® HIS (INR)	2.32 – 2.92	2.70	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Christoph Enne
Christoph ENNE

16-NOV-2015
Datum / Date:

**Freigabe von
Approved by**

Lucas Rosenmayr
Lucas ROSENMAYR

16. NOV. 2015
Datum / Date:
