
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ.02
---	---	-------------

Ceveron® Coagulation Control AK 5x1mL

REF 5011030
LOT 2B35B00.04
 31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Daptin TC	43.3 – 53.5	45.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

01. MRZ. 2016
Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

01. MRZ. 2016
Datum / Date: