



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2B00AZ.02

Coagulation Control AK 5x1mL

REF 5011050

LOT 2B35B00.02

31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	aPTT, Dapttin (sec)	43.3 – 53.5 sec	47.7	OK
	TECHNOPLASTIN® HIS (INR)	2.32 – 2.92	2.78	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lavdrim NASUFI

02-Jul-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

02-Jul-2014
Datum / Date: