
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1P00AZ.01
---	---	-------------

Ceveron® Coagulation Control N 5x1mL

REF 5020020
LOT 1P37B00.01
 30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LS (sec)	24.9 – 30.7 sec	25.8	OK
	FVIII Siron IP-Verd. (%)	65.3 – 108.8 %	88.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas ROSENMAYR

08-Apr-2015
Datum / Date:

**Freigabe von
Approved by**

Alessandro BENIGNI

08-Apr-2015
Datum / Date: