
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1P00AZ.02
-----------------------------------------------------------------------------------	-------------------------------------------------------	-------------

Coagulation Control N 5x1mL

REF 5020040
LOT 1P14B00.05
 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	FXI clotting	76 – 126 %	110.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Alessandro BENIGNI

05-Feb-2014

Datum / Date:

**Freigabe von
Approved by**

Lucas HOLOVICZ

05-Feb-2014

Datum / Date: