



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1P00AZ.01

Coagulation Control N 5x1mL

REF 5020040  
LOT 1P37B00.02  
⌚ 30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LS (sec.)	24.9 – 30.7	28.4	OK
	FII Technoplastin HIS (%)	73.5 – 122.5	94.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Monika WEINBAUER

10-Apr-2015  
Datum / Date:

Freigabe von  
Approved by

Lavdrim NASUFI

10-Apr-2015  
Datum / Date: