
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ1P00AZ.02
---	---	-------------

### Coagulation Control N 5x1ml

REF      5020040  
LOT      1P44B00.01  
      31-May-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	DAPTIN TC	32.0 – 39.4sec	36.3	OK
---	-----------	----------------	------	----

**Durchführung von**  
**Prepared by**

  
Alessandro BENIGNI

13-May-2014

Datum / Date:

**Freigabe von**  
**Approved by**

  
Lavdrim NASUFI

13-May-2014

Datum / Date: