



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1P00AZ.01

Coagulation Control N 50x1mL

REF 5020050
LOT 1P37B00.01
⌚ 30-Sep-2017

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|----------------------------|----------------------------------|----------------------------------|----|
| pH | | 7.4 – 7.8 | 7.6 | OK |
| Virologie Virology | HbsAg HIV 1,2 Ab HCV | Negative Negative Negative | Negative Negative Negative | OK |

| | | | | |
|--|--------------------------|------------------|------|----|
| Verwendbarkeit Usability Ceveron [®] alpha | Siron LS (sec.) | 24.9 – 30.7 sec. | 25.9 | OK |
| | FVIII Siron IP-Verd. (%) | 65.3 – 108.8 % | 88.7 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

Durchführung von
Prepared by


Lucas ROSENMAYR

08-Apr-2015
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

08-Apr-2015
Datum / Date: