
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1P00AZ.01
---	---	-------------

Coagulation Control N 50x1mL

REF 5020050
LOT 1P37B00.02
 30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron [®] alpha	F IX clotting	80.3 – 133.8 %	84.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUELI

09-Sep-2015
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

09-Sep-2015
Datum / Date: