



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3P00AZ.03

Ceveron Coagulation Control A 5x1ml

REF 5021035
LOT 3P38B00.02
⌚ 30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron [®] alpha	Siron LS	62.3 – 76.9	65.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

10-Dec-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

10-Dec-2014
Datum / Date: