

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3P00AZ.03
---	---	-------------

Coagulation Control A 5x1ml

REF 5021055
LOT 3P24B00.01
 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.15	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron [®] alpha	DAPTTIN TC, sec	88.5 – 119.5	110.5	OK
	FVIII Activity, (%)	11.9 – 19.9	16.3	OK

Durchführung von **Lavdrim NASUFI**
Prepared by



14-Jun-2013
Datum / Date:

Freigabe von **Harald SEVCIK**
Approved by



17. JUNI 2013
Datum / Date: