



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3P00AZ.03

Coagulation Control A 5x1ml

REF 5021055
LOT 3P24B00.03
⌚ 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.15	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, g/L	1.1 – 1.6	1.20	OK

Durchführung von
Prepared by

Alessandro BENIGNI

23-Jul-2013

Datum / Date:

Freigabe von
Approved by

Harald SEVCIK

23. JULI 2013

Datum / Date: