



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3P00AZ.03

Coagulation Control A 5x1mL

REF 5021055

LOT 3P24B00.04

30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.15	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Dapttin TC	88.5 – 119.7sec	105.5	OK
	FVIII chromogenes Substrat FVIII chromogenic substrate	14.0 – 23.4%	17.1	OK

Durchführung von  
Prepared by

Lavdrim NASUFI

29-Jun-2014  
Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI

29-JUN-2014  
Datum / Date: