




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3P00AZ.03

Coagulation Control A 5x1ml

REF 5021055

LOT 3P24B00.07

 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.2	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron [®] alpha	FXI clotting	27.0 – 45.0%	43.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lavdrim NASUFI

25-Mar-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOWICZ

25-Mar-2014
Datum / Date: