
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3P00AZ.03
---	---	-------------

**Coagulation Control A 5x1ml**

REF      5021055  
LOT      3P38B00.02  
      30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Siron LS, sec	62.3 – 76.9sec	68.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

Alessandro BENIGNI

22-Oct-2014  
Datum / Date:

**Freigabe von  
Approved by**

Lavdrim NASUFI

22-Oct-2014  
Datum / Date: