
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6V0AAZ.02
---	---	-------------

Dapttin TC 5x2mL

REF 5035060
LOT 6V45BA0.01
 31-Jul-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		6.6 – 7.2	6.9	OK	
Verwendbarkeit Usability Ceveron [®] alpha	Dapttin TC, sec	CCN LOT 1P44B00	32.0 – 39.4 sec	35.1	OK
		CCA LOT 3P0AB00	79.0 – 97.0 sec	91.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Alessandro BENIGNI

19-Aug-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

19-AUG-2014
Datum / Date: