

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6V0AAZ.02
---	---	-------------

Dapttin TC 5x2mL

REF 5035060
LOT 6V48BA0.01
 31-Oct-2016


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		6.6 – 7.2	6.9	OK	
Verwendbarkeit Usability Ceveron® alpha	Dapttin TC, sec	TCN LOT 1A31B02	32.7 – 40.3 sec	35.5	OK
		CCA LOT 3P38B00	89.1 – 110.1 sec	96.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by


Monika WEINBAUER

06-Nov-2015
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

06-Nov-2015
Datum / Date: