

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6V0AAZ.03
---	---	-------------

### Dapttin TC 5x2mL

REF      5035060  
LOT      6V54BA0.01  
      30-Sep-2017


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result			
pH	6.6 – 7.2	6.9	OK		
Verwendbarkeit Usability Ceveron® alpha	Dapttin TC, sec	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	33.1 – 40.9 sec	38.9	OK
		CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	96.5 – 119.1 sec	111.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

**Durchführung von  
Prepared by**

  
Lucas HOLOVICZ

29. FEB. 2016  
Datum / Date:

**Freigabe von  
Approved by**

  
Monika WEINBAUER

29. FEB. 2016  
Datum / Date: