



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6V0BAZ.02

Dapttin TC 6x10mL

REF 5035090  
LOT 6V47BB0.02  
⌚ 30-Sep-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		6.6 – 7.2	6.9	OK	
Verwendbarkeit Usability Ceveron® alpha	aPTT, sec	CCN LOT 1P44B00	32.0 – 39.4 sec	37.2	OK
		CCA LOT 3P38B00	84.7 – 114.5 sec	113.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Alessandro BENIGNI

13-Jan-2015  
Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI

13-Jan-2015  
Datum / Date: