
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6V0BAZ.02
---	---	-------------

Dapttin TC 6x10mL

REF 5035090
LOT 6V52BB0.01
 28-Feb-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH	6.6 – 7.2	6.9	OK		
Verwendbarkeit Usability Ceveron® alpha	aPTT, sec	CCN LOT 1P37B00	31.5 – 42.6 sec	37.1	OK
		CCA LOT 3P38B00	91.6 – 124.0 sec	105.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Fikret NASUFI

29-Jul-2015
Datum / Date:

Freigabe von
Approved by

[Signature]

29-Jul-2015
Datum / Date: