
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ6V0BAZ.02
---	---	-------------


### Daptin TC 20x10mL

REF      5035100  
LOT      6V36BB0.01  
      31-Aug-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	6.6 – 7.2	6.9	OK
Verwendbarkeit Usability Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P14B00	28.0 – 38.0 sec	34.1    OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P24B00	85.5 – 115.7 sec	103.8    OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von  
Prepared by


**Lavdrim NASUFI**

  
\_\_\_\_\_

22-Nov-2013  
Datum / Date:

Freigabe von  
Approved by

**Nikolaus BINDER**

  
\_\_\_\_\_

27-Nov-2013  
Datum / Date: