



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ2600AZ.01.

TECHNOVIEW ARIXTRA® Calibrtor Set


Technoclone

REF 50 900 10

LOT 2621B00

Herstelldatum: 09.08.2012

 2016 – 06 – 30

Inhalt	Reagenz	LOT	
1 x 1 ml	Technoview ARIXTRA® Cal 1	2101B02	30.06.2016
1 x 1 ml	Technoview ARIXTRA® Cal 2	2201B02	30.06.2016
1 x 1 ml	Technoview ARIXTRA® Cal 3	2301B02	30.06.2016
1 x 1 ml	Technoview ARIXTRA® Cal 4	2401B02	30.06.2016
1 x 1 ml	Technoview ARIXTRA® Cal 5	2501B02	30.06.2016

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	ARIXTRA® CON Low Lot 3U01B01	0,50 µg/ml (0,40 – 0,60)	0,48 µg/ml
	ARIXTRA® CON High Lot 3V01B01	1,48 µg/ml (1,29 – 1,67)	1,50 µg/ml
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes

Erstellt / created:

Unterschrift / signature: 

Datum / date: 09.08.12

Charge freigegeben / batch is released: Ja / yes

Nein / no

Freigegeben / released:

Unterschrift / signature: 

Datum / date: 9.8.12